

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000239  
11:16 20/07/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:33  
PRESSURIZING 3:36  
STERILIZING 4:56  
136.5 C 222kPa 4:56  
136.2 C 223kPa 5:26  
136.3 C 221kPa 5:56  
136.4 C 225kPa 6:26  
136.4 C 225kPa 6:56  
136.2 C 223kPa 7:26  
136.2 C 222kPa 7:56  
136.4 C 221kPa 8:26  
136.4 C 221kPa 8:26  
VENTING 8:26  
AIR DRYING 9:21  
CYCLE COMPLETE 9:29

-----  
Digital Signature #  
838FED89CE27B888  
-----