

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001795  
8:43 17/06/2021

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:45  
PRESSURIZING 4:53  
STERILIZING 5:56  
136.0 C 221kPa 5:56  
136.4 C 223kPa 6:26  
136.2 C 224kPa 6:56  
136.0 C 223kPa 7:26  
136.4 C 225kPa 7:56  
136.1 C 219kPa 8:26  
136.3 C 224kPa 8:56  
136.4 C 224kPa 9:26  
136.4 C 224kPa 9:26  
VENTING 9:26  
AIR DRYING 10:17  
CYCLE COMPLETE 13:21

-----

Digital Signature #  
2FD81477599DC054  
-----