
Midmark M11 - v1.0.0
Total Cycles: 4933

Name: RITTER 1
Sterilizer ID: V2178198
09/24/2024 08:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.4 F 0.0
2:00 72.0 F 0.0
4:00 79.5 F 0.0
6:00 112.1 F 0.1
8:00 141.1 F 0.1
10:00 173.5 F 0.3
12:00 193.1 F 2.7
14:00 216.6 F 6.3
16:00 220.9 F 9.7
18:00 231.3 F 13.5
20:00 232.2 F 8.3
22:00 243.2 F 13.1
24:00 240.5 F 10.6
26:00 236.1 F 8.5
28:00 250.5 F 15.2
30:00 261.8 F 21.6
32:00 272.2 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.7 F 28.7
2:00 271.7 F 27.9
3:00 272.4 F 28.7
4:00 272.3 F 28.4

Min 271.5 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.1 F 0.0
5:00 206.1 F 0.0
10:00 210.7 F 0.0
15:00 183.9 F 0.0
20:00 181.4 F 0.0
25:00 179.5 F 0.0
30:00 181.1 F 0.0

Drying Complete

FILLING: 0:43
HEATING: 32:32
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:10:40
CYCLE COMPLETE

09/24/2024 09:26 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____