
Midmark M11 - v1.0.0
Total Cycles: 4553

Name: RITTER 1
Sterilizer ID: V2178198
05/14/2024 03:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 119.6 F 0.0
2:00 124.1 F 1.6
4:00 149.7 F 4.6
6:00 174.8 F 8.6
8:00 193.4 F 13.3
10:00 216.0 F 6.8
12:00 226.9 F 11.7
14:00 230.8 F 7.4
16:00 242.0 F 13.2
18:00 241.5 F 11.1
20:00 253.7 F 17.5
22:00 265.2 F 24.2
24:00 272.0 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 271.7 F 28.9
2:00 271.7 F 28.9
3:00 271.7 F 28.9
4:00 271.6 F 28.8

Min 271.4 F 28.7
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.6 F 0.0
5:00 248.9 F 0.0
10:00 202.1 F 0.0
15:00 220.7 F 0.0
20:00 231.9 F 0.0
25:00 212.4 F 0.0
30:00 230.6 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 24:25
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:02:49
CYCLE COMPLETE

05/14/2024 04:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____