
Midmark M11 - v1.0.0
Total Cycles: 5098

Name: RITTER 1
Sterilizer ID: V2178198
11/15/2024 12:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.6 F 0.0
2:00 178.8 F 0.2
4:00 178.9 F 0.8
6:00 201.7 F 4.1
8:00 217.5 F 7.8
10:00 231.2 F 11.9
12:00 230.2 F 7.8
14:00 242.1 F 13.2
16:00 241.3 F 11.4
18:00 239.8 F 10.3
20:00 253.1 F 16.8
22:00 264.3 F 23.4
24:00 271.4 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.6 F 28.9
2:00 271.9 F 28.3
3:00 271.5 F 28.3
4:00 272.5 F 28.9

Min 271.3 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.7 F 0.0
5:00 223.9 F 0.0
10:00 224.7 F 0.0
15:00 229.7 F 0.0
20:00 225.6 F 0.0
25:00 225.5 F 0.0
30:00 230.5 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 24:04
STERILIZING: 4:00
VENTING: 3:53
DRYING: 30:00
TOTAL CYCLE: 01:02:56
CYCLE COMPLETE

11/15/2024 01:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____