
Midmark M11 - v1.0.0
Total Cycles: 1050

Name: RITTER 1
Sterilizer ID: V2178198
07/26/2021 07:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.2 F 0.0
2:00 169.0 F 1.7
4:00 182.9 F 5.1
6:00 202.4 F 9.9
8:00 218.5 F 6.3
10:00 230.7 F 8.8
12:00 240.8 F 12.0
14:00 240.5 F 11.1
16:00 237.4 F 9.2
18:00 251.1 F 15.8
20:00 262.4 F 22.2
22:00 272.7 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.7 F 28.8
2:00 271.3 F 28.2
3:00 272.7 F 29.0
4:00 271.6 F 28.2

Min 271.3 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.1 F 0.0
5:00 196.0 F 0.0
10:00 191.5 F 0.0
15:00 186.8 F 0.0
20:00 175.6 F 0.0
25:00 204.3 F 0.0
30:00 217.0 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 22:25
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:01:04
CYCLE COMPLETE

07/26/2021 09:07 AM
ACCEPTED: X
REJECTED:
Operator: U01