
Midmark M11 - v1.0.5
Total Cycles: 871

Name: RITTER2
Sterilizer ID: V2313228
03/10/2022 11:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.6 F 0.0
2:00 163.4 F 2.8
4:00 175.8 F 7.1
6:00 198.6 F 12.3
8:00 219.5 F 7.0
10:00 232.3 F 12.4
12:00 236.0 F 10.1
14:00 235.7 F 8.7
16:00 249.5 F 15.6
18:00 261.6 F 22.5
20:00 272.7 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.3 F 29.6
2:00 271.2 F 29.1
3:00 272.4 F 29.7
4:00 271.7 F 29.1

Min 271.2 F 28.9
Max 272.5 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.8 F 0.0
5:00 202.1 F 0.0
10:00 206.2 F 0.0
15:00 187.3 F 0.0
20:00 178.9 F 0.0
25:00 186.6 F 0.0
30:00 185.9 F 0.0
Drying Complete

FILLING: 1:44
HEATING: 20:24
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:17
CYCLE COMPLETE

03/10/2022 12:18 PM
ACCEPTED: X
REJECTED:
Operator: U02