

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3114

Name: RITTER 1  
Sterilizer ID: V2178198  
04/14/2023 11:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 140.1 F 0.0  
2:00 141.5 F 0.0  
4:00 158.7 F 2.7  
6:00 190.0 F 6.9  
8:00 218.1 F 12.1  
10:00 223.7 F 6.1  
12:00 235.0 F 10.7  
14:00 235.6 F 7.4  
16:00 243.8 F 12.9  
18:00 241.8 F 11.2  
20:00 254.4 F 17.6  
22:00 265.5 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.5 F 28.9  
2:00 271.4 F 28.3  
3:00 272.5 F 29.0  
4:00 271.6 F 28.3

Min 271.3 F 28.2  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.8 F 0.0  
5:00 212.9 F 0.0  
10:00 227.7 F 0.0  
15:00 232.5 F 0.0  
20:00 228.1 F 0.0  
25:00 230.4 F 0.0  
30:00 229.4 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 23:48  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:01:53  
CYCLE COMPLETE

04/14/2023 12:45 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_