

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3606

Name: RITTER2  
Sterilizer ID: V2313228  
05/10/2024 11:51 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 131.6 F 0.0  
2:00 131.8 F 1.5  
4:00 149.1 F 4.1  
6:00 175.7 F 7.5  
8:00 196.9 F 11.3  
10:00 222.7 F 5.3  
12:00 224.1 F 9.8  
14:00 233.5 F 9.2  
16:00 234.3 F 11.1  
18:00 234.5 F 8.8  
20:00 247.7 F 14.9  
22:00 258.6 F 21.0  
24:00 268.3 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.5  
1:00 271.9 F 29.9  
2:00 272.2 F 29.7  
3:00 272.8 F 30.4  
4:00 272.1 F 30.1

Min 270.8 F 29.4  
Max 273.6 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 170.0 F 0.0  
5:00 222.2 F 0.0  
10:00 184.1 F 0.0  
15:00 181.4 F 0.0  
20:00 179.8 F 0.0  
25:00 178.5 F 0.0  
30:00 177.0 F 0.0  
Drying Complete

FILLING: 1:00  
HEATING: 25:00  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:03:33  
CYCLE COMPLETE

05/10/2024 12:54 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_