

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 007187  
7:15 11/04/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:39  
PRESSURIZING 2:41  
STERILIZING 3:49  
136.4 C 221kPa 3:49  
136.2 C 221kPa 4:19  
136.1 C 220kPa 4:49  
136.3 C 218kPa 5:19  
136.4 C 218kPa 5:49  
136.2 C 222kPa 6:19  
136.4 C 224kPa 6:49  
136.2 C 223kPa 7:19  
136.2 C 223kPa 7:19  
VENTING 7:19  
AIR DRYING 8:09  
CYCLE COMPLETE 26:20

-----  
Digital Signature #  
2A4C84D3C32CB4B4  
-----