

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2865

Name: RITTER2  
Sterilizer ID: V2313228  
09/11/2023 05:30 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.0 F 0.0  
2:00 183.8 F 4.6  
4:00 208.0 F 10.5  
6:00 223.8 F 6.6  
8:00 238.3 F 12.9  
10:00 239.4 F 11.1  
12:00 238.9 F 10.3  
14:00 253.4 F 17.5  
16:00 265.7 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.8  
1:00 272.5 F 29.5  
2:00 271.5 F 29.0  
3:00 272.2 F 29.2  
4:00 271.7 F 29.2  
  
Min 271.4 F 28.7  
Max 272.6 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.7 F 0.0  
5:00 209.5 F 0.0  
10:00 221.8 F 0.0  
15:00 214.7 F 0.0  
20:00 218.3 F 0.0  
25:00 220.7 F 0.0  
30:00 219.4 F 0.0  
Drying Complete

FILLING: 2:51  
HEATING: 17:29  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 00:57:39  
CYCLE COMPLETE

09/11/2023 06:27 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_