

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000813  
15:06 01/12/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:53  
PRESSURIZING 2:59  
STERILIZING 4:02  
136.0 C 220kPa 4:02  
135.9 C 229kPa 4:32  
136.4 C 233kPa 5:02  
136.4 C 233kPa 5:32  
136.4 C 233kPa 6:02  
136.3 C 231kPa 6:32  
136.5 C 233kPa 7:02  
136.4 C 233kPa 7:32  
136.4 C 233kPa 7:32  
VENTING 7:32  
AIR DRYING 8:22  
CYCLE COMPLETE 23:22  
-----

Digital Signature #  
A2260310125A016F  
-----