
Midmark M11 - v1.0.0
Total Cycles: 1813

Name: RITTER 1
Sterilizer ID: V2178198
04/12/2022 08:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.4 F 0.0
2:00 74.6 F 0.0
4:00 78.7 F 0.0
6:00 94.9 F 0.1
8:00 128.7 F 1.7
10:00 162.4 F 5.0
12:00 184.2 F 8.2
14:00 223.1 F 11.6
16:00 218.3 F 3.7
18:00 229.4 F 7.8
20:00 238.5 F 11.9
22:00 234.3 F 8.1
24:00 246.3 F 12.8
26:00 244.0 F 12.0
28:00 255.9 F 18.2
30:00 266.4 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.6 F 28.7
2:00 271.5 F 27.9
3:00 272.6 F 28.9
4:00 272.1 F 28.3

Min 271.3 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.6 F 0.0
5:00 247.5 F 0.0
10:00 214.1 F 0.0
15:00 184.3 F 0.0
20:00 202.4 F 0.0
25:00 190.7 F 0.0
30:00 198.1 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 31:41
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:09:38
CYCLE COMPLETE

04/12/2022 09:37 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____