

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1125

Name: RITTER2  
Sterilizer ID: V2313228  
05/26/2022 02:33 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 170.6 F 0.0  
2:00 172.2 F 3.4  
4:00 192.8 F 8.3  
6:00 212.6 F 10.5  
8:00 226.1 F 10.2  
10:00 229.6 F 7.8  
12:00 239.7 F 11.1  
14:00 245.5 F 13.9  
16:00 258.3 F 20.9  
18:00 270.3 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.3  
1:00 272.4 F 30.5  
2:00 271.9 F 30.2  
3:00 272.0 F 29.8  
4:00 272.2 F 30.2  
  
Min 270.8 F 29.1  
Max 273.6 F 30.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.2 F 0.0  
5:00 226.7 F 0.0  
10:00 233.1 F 0.0  
15:00 230.0 F 0.0  
20:00 229.5 F 0.0  
25:00 228.9 F 0.0  
30:00 227.1 F 0.0  
Drying Complete

FILLING: 2:34  
HEATING: 18:45  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 00:58:23  
CYCLE COMPLETE

05/26/2022 03:31 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_