
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 003100
15:11 07/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:46
PRESSURIZING 2:47
STERILIZING 3:46
136.4 C 220kPa 3:46
136.4 C 225kPa 4:16
136.2 C 222kPa 4:46
136.4 C 225kPa 5:16
136.3 C 222kPa 5:46
136.4 C 219kPa 6:16
136.4 C 223kPa 6:46
136.4 C 226kPa 7:16
136.4 C 226kPa 7:16
VENTING 7:16
AIR DRYING 8:11
CYCLE COMPLETE 8:14

Digital Signature #
40EB33F493E95FD2
