
Midmark M11 - v1.0.0
Total Cycles: 3464

Name: RITTER 1
Sterilizer ID: V2178198
06/29/2023 09:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.0 F 0.0
2:00 167.0 F 2.6
4:00 182.7 F 6.6
6:00 204.3 F 11.5
8:00 219.3 F 6.1
10:00 233.4 F 11.6
12:00 233.8 F 8.6
14:00 239.9 F 10.3
16:00 246.3 F 13.5
18:00 258.7 F 20.2
20:00 269.8 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.5 F 29.2
2:00 272.0 F 28.6
3:00 271.5 F 28.6
4:00 272.5 F 29.2

Min 271.4 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.8 F 0.0
5:00 226.7 F 0.0
10:00 231.4 F 0.0
15:00 226.6 F 0.0
20:00 227.2 F 0.0
25:00 226.9 F 0.0
30:00 228.0 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 21:00
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:24
CYCLE COMPLETE

06/29/2023 10:54 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____