
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001451
12:49 09/24/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:45
PRESSURIZING 6:27
STERILIZING 7:58
136.0 C 220kPa 7:58
135.5 C 218kPa 8:28
136.2 C 219kPa 8:58
136.3 C 223kPa 9:28
136.3 C 219kPa 9:58
136.4 C 221kPa 10:28
136.3 C 222kPa 10:58
136.4 C 220kPa 11:28
VENTING 11:28
AIR DRYING 12:31
CYCLE COMPLETE 19:15

Digital Signature #
DA106A9FDC75C31A
