
Midmark M11 - v1.0.0
Total Cycles: 5001

Name: RITTER 1
Sterilizer ID: V2178198
10/16/2024 03:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.0 F 0.0
2:00 134.2 F 0.2
4:00 152.9 F 0.9
6:00 183.4 F 3.8
8:00 201.8 F 7.8
10:00 229.6 F 12.0
12:00 226.1 F 6.5
14:00 235.7 F 11.2
16:00 233.3 F 7.7
18:00 246.3 F 13.4
20:00 243.5 F 11.9
22:00 255.2 F 17.9
24:00 265.9 F 24.3
26:00 271.8 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.5
1:00 271.8 F 28.8
2:00 272.6 F 29.3
3:00 272.6 F 28.9
4:00 271.7 F 28.5

Min 271.3 F 28.3
Max 273.2 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.0 F 0.0
5:00 210.6 F 0.0
10:00 198.9 F 0.0
15:00 230.3 F 0.0
20:00 247.2 F 0.0
25:00 243.1 F 0.0
30:00 233.1 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 26:30
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:05:09
CYCLE COMPLETE

10/16/2024 04:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____