
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007988
9:09 23/10/2024

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:18
PRESSURIZING 3:22
STERILIZING 4:23
136.2 C 220kPa 4:23
136.4 C 224kPa 4:53
136.4 C 223kPa 5:23
136.2 C 219kPa 5:53
136.4 C 219kPa 6:23
136.4 C 223kPa 6:53
136.4 C 223kPa 7:23
136.4 C 223kPa 7:53
136.4 C 223kPa 7:53
VENTING 7:53
AIR DRYING 8:47
CYCLE COMPLETE 9:31

Digital Signature #
50A94E672883C044
