

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1385

Name: RITTER 1  
Sterilizer ID: V2178198  
11/18/2021 12:45 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 208.0 F 0.0  
2:00 193.7 F 3.7  
4:00 212.3 F 9.5  
6:00 224.0 F 7.0  
8:00 236.1 F 11.2  
10:00 235.3 F 8.9  
12:00 238.5 F 9.5  
14:00 249.8 F 14.9  
16:00 262.4 F 22.0  
18:00 272.3 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.0  
1:00 272.3 F 28.5  
2:00 271.7 F 28.4  
3:00 272.2 F 28.4  
4:00 271.5 F 28.2  
  
Min 271.3 F 27.9  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.1 F 0.0  
5:00 231.5 F 0.0  
10:00 217.1 F 0.0  
15:00 228.8 F 0.0  
20:00 228.4 F 0.0  
25:00 229.8 F 0.0  
30:00 221.9 F 0.0  
Drying Complete

FILLING: 2:58  
HEATING: 18:12  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:20  
CYCLE COMPLETE

11/18/2021 01:43 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_