
Midmark M11 - v1.0.5
Total Cycles: 595

Name: RITTER2
Sterilizer ID: V2313228
12/01/2021 12:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.2 F 0.0
2:00 149.1 F 2.5
4:00 163.7 F 6.1
6:00 189.3 F 11.2
8:00 212.8 F 4.8
10:00 225.9 F 10.1
12:00 230.6 F 7.9
14:00 242.8 F 13.0
16:00 242.8 F 12.0
18:00 255.8 F 18.8
20:00 267.2 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 271.6 F 28.8
2:00 271.7 F 28.8
3:00 271.9 F 29.0
4:00 272.4 F 29.3

Min 271.5 F 28.8
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.0 F 0.0
5:00 228.1 F 0.0
10:00 215.9 F 0.0
15:00 216.1 F 0.0
20:00 224.5 F 0.0
25:00 222.1 F 0.0
30:00 220.3 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 21:18
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:59:48
CYCLE COMPLETE

12/01/2021 02:19 PM
ACCEPTED: X
REJECTED:
Operator: U02