
Midmark M11 - v1.0.0
Total Cycles: 3418

Name: RITTER 1
Sterilizer ID: V2178198
06/21/2023 07:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.5 F 0.0
2:00 75.0 F 0.0
4:00 86.9 F 0.0
6:00 112.2 F 0.0
8:00 147.4 F 1.2
10:00 173.8 F 3.9
12:00 189.6 F 7.1
14:00 206.0 F 10.7
16:00 219.2 F 6.4
18:00 225.1 F 7.7
20:00 235.5 F 11.9
22:00 234.6 F 8.1
24:00 244.4 F 12.4
26:00 244.0 F 12.0
28:00 255.6 F 18.0
30:00 265.9 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.8 F 28.9
2:00 271.5 F 28.0
3:00 272.5 F 29.0
4:00 272.1 F 28.4

Min 271.3 F 28.0
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.9 F 0.0
5:00 217.1 F 0.0
10:00 233.7 F 0.0
15:00 233.6 F 0.0
20:00 224.9 F 0.0
25:00 222.8 F 0.0
30:00 223.0 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:50
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:09:46
CYCLE COMPLETE

06/21/2023 08:27 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____