
Midmark M11 - v1.0.5
Total Cycles: 1876

Name: RITTER2
Sterilizer ID: V2313228
12/21/2022 05:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.8 F 0.0
2:00 171.5 F 3.6
4:00 189.0 F 8.1
6:00 209.5 F 13.0
8:00 224.1 F 8.0
10:00 235.6 F 12.5
12:00 239.2 F 11.4
14:00 238.8 F 10.3
16:00 253.0 F 17.3
18:00 265.0 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.4 F 29.5
2:00 271.5 F 29.2
3:00 272.4 F 29.6
4:00 271.6 F 28.9

Min 271.4 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.6 F 0.0
5:00 227.8 F 0.0
10:00 238.6 F 0.0
15:00 231.2 F 0.0
20:00 226.9 F 0.0
25:00 227.6 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 1:53
HEATING: 19:47
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

12/21/2022 06:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____