
Midmark M11 - v1.0.0
Total Cycles: 661

Name: MIDMARK
Sterilizer ID: V2178198
02/10/2021 12:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.3 F 0.0
2:00 128.7 F 1.2
4:00 154.2 F 4.3
6:00 182.4 F 8.9
8:00 229.0 F 7.6
10:00 224.8 F 9.3
12:00 235.0 F 8.5
14:00 241.7 F 12.3
16:00 241.8 F 10.8
18:00 255.6 F 18.0
20:00 268.2 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.9 F 28.7
2:00 272.1 F 28.8
3:00 271.5 F 28.1
4:00 272.8 F 28.8

Min 271.1 F 28.1
Max 273.1 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.3 F 0.1
5:00 242.2 F 0.0
10:00 230.4 F 0.0
15:00 233.7 F 0.0
20:00 225.0 F 0.0
Dry Cycle Aborted

FILLING: 0:58
HEATING: 21:03
STERILIZING: 4:00
VENTING: 2:56
DRYING: 20:57
TOTAL CYCLE: 00:49:55

02/10/2021 12:56 PM
ACCEPTED: X
REJECTED:
Operator: U01