
Midmark M11 - v1.0.5
Total Cycles: 1559

Name: RITTER2
Sterilizer ID: V2313228
10/04/2022 10:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.7 F 0.0
2:00 188.7 F 0.1
4:00 202.8 F 2.6
6:00 219.5 F 7.7
8:00 235.9 F 13.1
10:00 236.5 F 10.4
12:00 236.5 F 8.8
14:00 239.9 F 9.9
16:00 250.3 F 15.6
18:00 262.6 F 22.7
20:00 272.3 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.6 F 29.2
2:00 271.6 F 28.6
3:00 272.7 F 29.3
4:00 271.8 F 28.6

Min 271.5 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.7 F 0.0
5:00 226.9 F 0.0
10:00 237.4 F 0.0
15:00 230.8 F 0.0
20:00 222.2 F 0.0
25:00 221.5 F 0.0
30:00 220.2 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 20:12
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:43
CYCLE COMPLETE

10/04/2022 11:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____