

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3022

Name: RITTER2  
Sterilizer ID: V2313228  
10/17/2023 04:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.9 F 0.0  
2:00 189.8 F 4.6  
4:00 216.3 F 11.5  
6:00 228.0 F 8.2  
8:00 237.3 F 9.8  
10:00 242.5 F 11.4  
12:00 249.0 F 14.9  
14:00 263.5 F 23.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.6  
1:00 272.1 F 28.7  
2:00 272.8 F 29.3  
3:00 271.8 F 28.8  
4:00 272.4 F 28.9  
  
Min 271.6 F 28.4  
Max 272.9 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.4 F 0.0  
5:00 219.2 F 0.0  
10:00 220.0 F 0.0  
15:00 222.1 F 0.0  
20:00 219.9 F 0.0  
25:00 217.4 F 0.0  
30:00 229.4 F 0.0  
  
Drying Complete

FILLING: 2:37  
HEATING: 15:45  
STERILIZING: 4:00  
VENTING: 2:59  
DRYING: 30:00  
TOTAL CYCLE: 00:55:21  
CYCLE COMPLETE

10/17/2023 05:10 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_