
Midmark M11 - v1.0.5
Total Cycles: 554

Name: RITTER2
Sterilizer ID: V2313228
11/16/2021 12:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.7 F 0.0
2:00 74.4 F 0.0
4:00 77.2 F 0.0
6:00 99.9 F 0.1
8:00 136.6 F 0.2
10:00 171.7 F 0.3
12:00 195.8 F 1.9
14:00 210.4 F 5.1
16:00 220.7 F 8.4
18:00 230.5 F 12.0
20:00 230.3 F 8.4
22:00 240.7 F 13.5
24:00 241.2 F 11.7
26:00 240.4 F 10.7
28:00 253.1 F 17.0
30:00 263.8 F 23.4
32:00 272.0 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.5 F 29.2
2:00 271.2 F 28.5
3:00 272.5 F 29.3
4:00 271.1 F 28.5

Min 271.0 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.2 F 0.0
5:00 234.9 F 0.0
10:00 219.7 F 0.0
15:00 218.9 F 0.0
20:00 227.5 F 0.0
25:00 230.7 F 0.0
30:00 229.2 F 0.0

Drying Complete

FILLING: 1:25
HEATING: 32:10
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:10:47
CYCLE COMPLETE

11/16/2021 01:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____