
Midmark M11 - v1.0.0
Total Cycles: 1756

Name: RITTER 1
Sterilizer ID: V2178198
03/28/2022 03:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.5 F 0.0
2:00 137.4 F 0.1
4:00 161.6 F 2.5
6:00 192.8 F 6.1
8:00 210.4 F 11.2
10:00 221.2 F 5.4
12:00 232.4 F 10.6
14:00 231.8 F 7.1
16:00 245.1 F 13.5
18:00 243.8 F 12.1
20:00 256.6 F 18.8
22:00 268.0 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.4 F 28.9
2:00 271.4 F 28.2
3:00 272.6 F 29.1
4:00 271.8 F 28.4

Min 271.4 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.7 F 0.0
5:00 229.2 F 0.0
10:00 206.7 F 0.0
15:00 202.9 F 0.0
20:00 228.6 F 0.0
25:00 233.3 F 0.0
30:00 226.4 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 23:17
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:01:28
CYCLE COMPLETE

03/28/2022 04:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____