
Midmark M11 - v1.0.0
Total Cycles: 1933

Name: RITTER 1
Sterilizer ID: V2178198
05/11/2022 02:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.9 F 0.0
2:00 201.0 F 4.0
4:00 202.2 F 9.0
6:00 221.2 F 8.0
8:00 230.6 F 10.0
10:00 233.0 F 7.9
12:00 244.6 F 13.1
14:00 243.4 F 11.8
16:00 256.3 F 18.6
18:00 267.7 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.1
1:00 272.6 F 28.9
2:00 271.5 F 28.0
3:00 272.6 F 28.8
4:00 271.5 F 28.0

Min 271.4 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 231.4 F 0.0
10:00 231.8 F 0.0
15:00 232.0 F 0.0
20:00 227.1 F 0.0
25:00 226.3 F 0.0
30:00 227.6 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 19:11
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:18
CYCLE COMPLETE

05/11/2022 03:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____