

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.9uS / 0.5ppm  
CYCLE NUMBER 000635  
11:53 08/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:18  
PRESSURIZING 3:27  
STERILIZING 4:38  
136.0 C 227kPa 4:38  
136.3 C 231kPa 5:08  
136.4 C 232kPa 5:38  
136.4 C 232kPa 6:08  
136.4 C 232kPa 6:38  
136.2 C 230kPa 7:08  
136.4 C 231kPa 7:38  
136.4 C 232kPa 8:08  
136.4 C 232kPa 8:08  
VENTING 8:08  
AIR DRYING 9:02  
CYCLE COMPLETE 24:02

-----  
Digital Signature #  
0D7B17078150D44D  
-----