
Midmark M11 - v1.0.0
Total Cycles: 3050

Name: RITTER 1
Sterilizer ID: V2178198
03/29/2023 07:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 80.9 F 0.0
2:00 82.4 F 0.0
4:00 113.6 F 0.0
6:00 124.7 F 0.0
8:00 163.0 F 0.1
10:00 187.9 F 2.8
12:00 203.4 F 6.2
14:00 216.0 F 9.5
16:00 227.9 F 13.3
18:00 230.1 F 8.2
20:00 240.3 F 12.9
22:00 239.4 F 10.2
24:00 237.4 F 8.2
26:00 249.7 F 14.9
28:00 261.4 F 21.5
30:00 272.2 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.5 F 28.7
2:00 271.3 F 28.0
3:00 272.6 F 28.9
4:00 271.7 F 28.1

Min 271.2 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.6 F 0.0
5:00 204.4 F 0.0
10:00 217.7 F 0.0
15:00 221.8 F 0.0
20:00 230.6 F 0.0
25:00 231.0 F 0.0
30:00 226.7 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 30:30
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:08:28
CYCLE COMPLETE

03/29/2023 08:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____