
Midmark M11 - v1.0.0
Total Cycles: 2715

Name: RITTER 1
Sterilizer ID: V2178198
12/22/2022 07:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 77.7 F 0.0
2:00 78.3 F 0.0
4:00 81.9 F 0.0
6:00 114.0 F 0.5
8:00 152.4 F 2.3
10:00 178.7 F 4.9
12:00 215.1 F 8.5
14:00 225.3 F 12.0
16:00 224.2 F 5.2
18:00 232.3 F 9.3
20:00 240.9 F 13.4
22:00 237.5 F 9.7
24:00 238.7 F 9.1
26:00 245.7 F 12.9
28:00 257.2 F 19.1
30:00 267.7 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 271.7 F 28.7
2:00 271.8 F 28.8
3:00 271.7 F 28.8
4:00 271.7 F 28.7

Min 271.6 F 28.5
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.2 F 0.0
5:00 230.2 F 0.0
10:00 220.3 F 0.0
15:00 183.6 F 0.0
20:00 170.8 F 0.0
25:00 190.7 F 0.0
30:00 207.7 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 31:59
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:10:02
CYCLE COMPLETE

12/22/2022 09:04 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____