
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001145
11:16 24/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:53
PRESSURIZING 2:59
STERILIZING 4:10
136.0 C 220kPa 4:10
136.2 C 224kPa 4:40
136.2 C 222kPa 5:10
136.4 C 225kPa 5:40
136.2 C 223kPa 6:10
136.1 C 223kPa 6:40
136.3 C 224kPa 7:10
136.4 C 223kPa 7:40
136.4 C 223kPa 7:40
VENTING 7:40
AIR DRYING 8:31
CYCLE COMPLETE 23:31

Digital Signature #
F3FB7A986F13ABA8
