
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000630
11:00 03/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:43
PRESSURIZING 2:50
STERILIZING 3:56
136.0 C 227kPa 3:56
136.1 C 231kPa 4:26
136.3 C 232kPa 4:56
136.4 C 233kPa 5:26
136.3 C 231kPa 5:56
136.3 C 232kPa 6:26
136.3 C 232kPa 6:56
136.4 C 230kPa 7:26
136.4 C 230kPa 7:26
VENTING 7:26
AIR DRYING 8:19
CYCLE COMPLETE 23:19

Digital Signature #
555361CE727DC0D3
