

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1444

Name: RITTER2  
Sterilizer ID: V2313228  
08/31/2022 05:52 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 192.9 F 0.0  
2:00 192.5 F 3.3  
4:00 207.2 F 8.7  
6:00 221.4 F 9.0  
8:00 231.9 F 11.1  
10:00 233.4 F 9.2  
12:00 236.7 F 9.4  
14:00 249.1 F 15.4  
16:00 261.6 F 22.6  
18:00 272.8 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 29.2  
1:00 272.3 F 29.7  
2:00 271.3 F 29.1  
3:00 272.3 F 29.9  
4:00 271.8 F 29.4  
  
Min 271.0 F 29.0  
Max 272.4 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.4 F 0.0  
5:00 231.6 F 0.0  
10:00 234.0 F 0.0  
15:00 229.0 F 0.0  
20:00 223.2 F 0.0  
25:00 226.5 F 0.0  
30:00 228.1 F 0.0  
Drying Complete

FILLING: 2:09  
HEATING: 18:22  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 30:00  
TOTAL CYCLE: 00:57:35  
CYCLE COMPLETE

08/31/2022 06:49 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_