

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3168

Name: RITTER 1  
Sterilizer ID: V2178198  
04/27/2023 11:14 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 146.6 F 0.0  
2:00 148.3 F 0.0  
4:00 173.2 F 2.1  
6:00 197.0 F 5.7  
8:00 210.3 F 10.2  
10:00 220.0 F 5.9  
12:00 232.1 F 9.4  
14:00 237.6 F 10.3  
16:00 242.9 F 12.1  
18:00 241.3 F 10.8  
20:00 254.3 F 17.4  
22:00 265.8 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.1  
1:00 272.3 F 28.5  
2:00 272.0 F 28.6  
3:00 271.9 F 28.3  
4:00 272.0 F 28.7

Min 271.4 F 28.0  
Max 272.6 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.9 F 0.0  
5:00 205.5 F 0.0  
10:00 220.3 F 0.0  
15:00 216.3 F 0.0  
20:00 213.1 F 0.0  
25:00 211.2 F 0.0  
30:00 210.4 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 23:39  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 01:01:58  
CYCLE COMPLETE

04/27/2023 12:15 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_