
Midmark M11 - v1.0.5
Total Cycles: 1715

Name: RITTER2
Sterilizer ID: V2313228
11/14/2022 10:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.7 F 0.0
2:00 169.7 F 0.1
4:00 178.2 F 0.6
6:00 207.6 F 3.8
8:00 223.2 F 8.9
10:00 235.0 F 10.3
12:00 238.2 F 11.1
14:00 237.2 F 9.6
16:00 238.6 F 9.1
18:00 251.7 F 16.4
20:00 264.2 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.6 F 29.3
2:00 271.6 F 28.6
3:00 272.6 F 29.5
4:00 272.1 F 28.9

Min 271.6 F 28.6
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 213.9 F 0.0
10:00 227.7 F 0.0
15:00 229.6 F 0.0
20:00 230.7 F 0.0
25:00 225.6 F 0.0
30:00 226.2 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 21:53
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:59:59
CYCLE COMPLETE

11/14/2022 11:58 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____