

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 001256  
9:29 15/09/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:29  
PRESSURIZING 3:32  
STERILIZING 4:59  
136.4 C 223kPa 4:59  
136.3 C 223kPa 5:29  
136.2 C 222kPa 5:59  
136.1 C 222kPa 6:29  
136.3 C 224kPa 6:59  
136.4 C 224kPa 7:29  
136.2 C 222kPa 7:59  
136.3 C 223kPa 8:29  
136.3 C 223kPa 8:29  
VENTING 8:29  
AIR DRYING 9:25  
CYCLE COMPLETE 24:25  
-----

Digital Signature #  
4CDCB8FC75E01F23  
-----