
Midmark M11 - v1.0.0
Total Cycles: 2340

Name: RITTER 1
Sterilizer ID: V2178198
09/20/2022 07:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.4 F 0.0
2:00 76.9 F 0.0
4:00 85.1 F 0.0
6:00 105.7 F 0.0
8:00 151.6 F 0.9
10:00 171.0 F 4.6
12:00 214.5 F 8.1
14:00 214.0 F 11.7
16:00 220.1 F 5.0
18:00 233.4 F 9.4
20:00 238.3 F 12.0
22:00 238.8 F 10.5
24:00 235.1 F 7.7
26:00 248.5 F 14.5
28:00 259.8 F 20.8
30:00 270.4 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 271.9 F 28.8
2:00 272.5 F 29.2
3:00 271.9 F 29.2
4:00 271.5 F 28.6

Min 271.2 F 28.6
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 242.2 F 0.0
10:00 228.9 F 0.0
15:00 226.1 F 0.0
20:00 224.5 F 0.0
25:00 221.7 F 0.0
30:00 220.8 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 31:30
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:09:34
CYCLE COMPLETE

09/20/2022 08:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____