
Midmark M11 - v1.0.5
Total Cycles: 4457

Name: RITTER2
Sterilizer ID: V2313228
03/26/2025 09:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.9 F 0.0
2:00 73.9 F 0.2
4:00 77.6 F 0.8
6:00 86.4 F 2.9
8:00 118.5 F 5.7
10:00 145.2 F 8.8
12:00 169.7 F 12.2
14:00 207.3 F 3.6
16:00 212.7 F 6.7
18:00 223.4 F 10.2
20:00 234.9 F 9.9
22:00 234.1 F 10.0
24:00 238.0 F 9.6
26:00 245.0 F 13.3
28:00 255.7 F 19.1
30:00 265.7 F 25.3
32:00 270.7 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 30.1
1:00 272.1 F 30.2
2:00 272.1 F 30.1
3:00 272.2 F 30.1
4:00 272.4 F 30.2

Min 271.9 F 30.0
Max 272.9 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.5 F 0.0
5:00 207.0 F 0.0
10:00 211.6 F 0.0
15:00 204.2 F 0.0
20:00 185.2 F 0.0
25:00 183.7 F 0.0
30:00 187.0 F 0.0

Drying Complete

FILLING: 1:07
HEATING: 32:38
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:11:32
CYCLE COMPLETE

03/26/2025 10:18 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____