
Midmark M11 - v1.0.5
Total Cycles: 3233

Name: RITTER2
Sterilizer ID: V2313228
12/19/2023 01:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.8 F 0.0
2:00 144.0 F 0.4
4:00 143.7 F 2.9
6:00 171.5 F 6.3
8:00 189.3 F 10.3
10:00 208.6 F 5.8
12:00 221.2 F 7.9
14:00 230.2 F 12.4
16:00 232.6 F 8.9
18:00 240.2 F 11.0
20:00 244.4 F 13.1
22:00 256.1 F 19.3
24:00 266.7 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.4 F 29.7
2:00 271.4 F 29.3
3:00 272.4 F 29.8
4:00 271.5 F 29.1

Min 271.3 F 29.0
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.7 F 0.0
5:00 209.3 F 0.0
10:00 192.5 F 0.0
15:00 189.3 F 0.0
20:00 207.0 F 0.0
25:00 214.2 F 0.0
30:00 221.8 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 25:35
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:03:56
CYCLE COMPLETE

12/19/2023 02:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____