

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4948

Name: RITTER 1  
Sterilizer ID: V2178198  
10/01/2024 01:20 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 132.3 F 0.0  
2:00 134.5 F 0.1  
4:00 147.9 F 0.4  
6:00 189.2 F 2.6  
8:00 201.4 F 6.5  
10:00 215.6 F 10.1  
12:00 230.8 F 8.9  
14:00 228.7 F 8.4  
16:00 239.2 F 12.9  
18:00 238.6 F 10.1  
20:00 237.1 F 8.8  
22:00 248.9 F 14.6  
24:00 260.2 F 20.8  
26:00 270.5 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.2  
1:00 272.2 F 28.9  
2:00 272.1 F 28.5  
3:00 271.2 F 28.0  
4:00 271.7 F 28.7

Min 271.0 F 28.0  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.1 F 0.0  
5:00 203.9 F 0.0  
10:00 198.5 F 0.0  
15:00 203.3 F 0.0  
20:00 228.5 F 0.0  
25:00 243.5 F 0.0  
30:00 238.7 F 0.0

Drying Complete

FILLING: 0:49  
HEATING: 26:55  
STERILIZING: 4:00  
VENTING: 3:52  
DRYING: 30:00  
TOTAL CYCLE: 01:05:36  
CYCLE COMPLETE

10/01/2024 02:25 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_