
Midmark M11 - v1.0.5
Total Cycles: 703

Name: RITTER2
Sterilizer ID: V2313228
01/19/2022 09:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.5 F 0.0
2:00 77.0 F 0.2
4:00 110.8 F 1.1
6:00 151.1 F 3.1
8:00 175.7 F 6.0
10:00 176.2 F 9.3
12:00 185.8 F 13.4
14:00 203.7 F 5.7
16:00 215.3 F 9.6
18:00 226.2 F 10.4
20:00 234.0 F 9.9
22:00 236.5 F 9.5
24:00 245.3 F 13.7
26:00 256.6 F 19.8
28:00 266.9 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 272.5 F 29.9
2:00 271.4 F 29.6
3:00 272.3 F 29.9
4:00 271.4 F 29.6

Min 271.3 F 29.3
Max 273.0 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.3 F 0.0
5:00 200.0 F 0.0
10:00 187.9 F 0.0
15:00 183.7 F 0.0
20:00 186.2 F 0.0
25:00 188.7 F 0.0
30:00 201.1 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 29:35
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:07:34
CYCLE COMPLETE

01/19/2022 10:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____