
Midmark M11 - v1.0.5
Total Cycles: 3193

Name: RITTER2
Sterilizer ID: V2313228
12/06/2023 04:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.2 F 0.0
2:00 142.1 F 1.7
4:00 166.8 F 5.2
6:00 189.0 F 9.7
8:00 213.0 F 6.9
10:00 223.7 F 8.7
12:00 235.0 F 11.9
14:00 237.5 F 11.0
16:00 236.4 F 9.3
18:00 250.2 F 15.9
20:00 261.6 F 22.5
22:00 272.2 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.5 F 29.9
2:00 272.3 F 29.4
3:00 271.2 F 29.0
4:00 272.1 F 29.9

Min 271.2 F 28.9
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.1 F 0.0
5:00 214.7 F 0.0
10:00 215.9 F 0.0
15:00 211.7 F 0.0
20:00 208.9 F 0.0
25:00 212.9 F 0.0
30:00 219.5 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 22:33
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:01:16
CYCLE COMPLETE

12/06/2023 05:52 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____