
Midmark M11 - v1.0.5
Total Cycles: 447

Name: RITTER2
Sterilizer ID: V2313228
10/14/2021 06:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.5 F 0.0
2:00 81.2 F 0.0
4:00 119.2 F 0.0
6:00 120.0 F 0.2
8:00 156.3 F 0.7
10:00 186.0 F 4.3
12:00 207.3 F 8.0
14:00 225.4 F 11.7
16:00 225.5 F 5.8
18:00 235.4 F 10.2
20:00 235.7 F 8.9
22:00 243.7 F 12.6
24:00 241.9 F 11.5
26:00 254.5 F 18.0
28:00 265.5 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.5 F 29.3
2:00 271.3 F 28.9
3:00 272.6 F 29.2
4:00 271.4 F 28.9

Min 271.2 F 28.5
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 217.2 F 0.0
5:00 226.1 F 0.0
10:00 187.4 F 0.0
15:00 190.8 F 0.0
20:00 195.9 F 0.0
25:00 204.8 F 0.0
Dry Cycle Aborted

FILLING: 1:04
HEATING: 29:46
STERILIZING: 4:00
VENTING: 3:01
DRYING: 26:26
TOTAL CYCLE: 01:04:17

10/14/2021 07:50 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____