
Midmark M11 - v1.0.0
Total Cycles: 3475

Name: RITTER 1
Sterilizer ID: V2178198
06/30/2023 02:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.6 F 0.0
2:00 159.7 F 2.6
4:00 178.1 F 6.2
6:00 198.8 F 10.9
8:00 218.2 F 5.0
10:00 231.3 F 10.1
12:00 234.5 F 8.0
14:00 243.2 F 12.8
16:00 241.7 F 11.2
18:00 254.6 F 17.8
20:00 265.8 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.4 F 28.9
2:00 271.3 F 28.4
3:00 272.4 F 29.1
4:00 271.7 F 28.4

Min 271.3 F 28.3
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.2 F 0.0
5:00 231.6 F 0.0
10:00 235.5 F 0.0
15:00 225.2 F 0.0
20:00 223.1 F 0.0
25:00 223.6 F 0.0
30:00 225.5 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 21:40
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:57
CYCLE COMPLETE

06/30/2023 03:44 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____