
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001050
11:48 29/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:35
STERILIZING 3:39
136.4 C 222kPa 3:39
136.1 C 222kPa 4:09
136.1 C 222kPa 4:39
136.0 C 222kPa 5:09
136.2 C 223kPa 5:39
136.4 C 225kPa 6:09
136.2 C 224kPa 6:39
136.3 C 223kPa 7:09
136.3 C 223kPa 7:09
VENTING 7:10
AIR DRYING 7:59
CYCLE COMPLETE 8:10

Digital Signature #
B2D43EA58848C29E
