
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002391
8:58 17/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:14
PRESSURIZING 3:16
STERILIZING 4:27
136.0 C 220kPa 4:27
136.0 C 223kPa 4:57
136.5 C 225kPa 5:27
136.1 C 221kPa 5:57
136.4 C 224kPa 6:27
136.3 C 223kPa 6:57
136.1 C 223kPa 7:27
136.2 C 223kPa 7:57
136.2 C 223kPa 7:57
VENTING 7:58
AIR DRYING 8:52
CYCLE COMPLETE 26:43

Digital Signature #
89EBDC628E675B48
