
Midmark M11 - v1.0.5
Total Cycles: 3478

Name: RITTER2
Sterilizer ID: V2313228
03/26/2024 02:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 86.5 F 0.0
2:00 87.7 F 0.0
4:00 105.5 F 0.2
6:00 132.2 F 2.2
8:00 156.3 F 5.0
10:00 172.2 F 8.2
12:00 189.9 F 11.7
14:00 212.7 F 4.7
16:00 221.4 F 8.6
18:00 233.7 F 12.9
20:00 234.4 F 9.6
22:00 238.6 F 10.2
24:00 244.9 F 13.4
26:00 256.1 F 19.3
28:00 266.2 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.1
1:00 271.6 F 29.6
2:00 272.5 F 29.7
3:00 271.3 F 29.0
4:00 271.8 F 29.9

Min 271.0 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.9 F 0.0
5:00 224.9 F 0.0
10:00 227.1 F 0.0
15:00 220.2 F 0.0
20:00 221.0 F 0.0
25:00 225.9 F 0.0
30:00 233.9 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 29:37
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:08:10
CYCLE COMPLETE

03/26/2024 03:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____