
Midmark M11 - v1.0.0
Total Cycles: 3666

Name: RITTER 1
Sterilizer ID: V2178198
08/16/2023 04:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.0 F 0.0
2:00 148.1 F 2.8
4:00 173.4 F 6.9
6:00 195.2 F 11.5
8:00 213.9 F 4.3
10:00 226.8 F 9.0
12:00 236.0 F 10.7
14:00 240.7 F 11.1
16:00 237.9 F 9.4
18:00 252.0 F 16.3
20:00 264.1 F 23.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.2
1:00 272.2 F 28.4
2:00 272.7 F 28.9
3:00 271.6 F 28.3
4:00 272.2 F 28.5

Min 271.4 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.3 F 0.0
5:00 201.1 F 0.0
10:00 209.8 F 0.0
15:00 213.0 F 0.0
20:00 209.2 F 0.0
25:00 207.5 F 0.0
30:00 204.6 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 21:54
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:18
CYCLE COMPLETE

08/16/2023 05:54 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____